



The National Physical Laboratory (NPL)


- Recent Nuclear Medicine Work
 - Incorporation of new electrometer system
 - Switching unit : Low leakage current
 - Attempt to minimise waiting periods
 - Calibration Factors for I-125 brachytherapy seeds
 - Investigation into commercial calibrator linearity
 - RCUF issues : will discuss tomorrow

I-125 Brachytherapy seeds

- New range of I-125 seeds
 - Various manufacturers (ie: BEBIG etc .., S06, S17)
 - Becoming more widespread in use



NPL Air kerma Primary Standard



Air KERMA rate measurements at NPL: Primary Standard : 50 kV free air chamber



NE 2551 protection level secondary standard



- Spherical Ion Chamber calibrated at NPL in terms of air kerma in appropriate fields
 - ISO 4037-1 narrow spectrum X-rays
 - Cs-137
 - Co-60
 - Response factor calculated
- 3 litres sensitive volume
- Sources placed in low density holder 1 m from floor in large room
- Sources rotated through 360 degrees, and mean response calculated (source uniformity)
- Temperature controlled room
 - Temperature and Pressure monitored throughout

Corrections to measurements

- Background/leakage current
- Air density (20 °C and 101.3 kPa)
- Air attenuation
- Radiation scatter
- Chamber size correction
- Radioactive decay

Summary of AKR measurements at NPL

Table 1 – Summary of NPL Calibrations in terms of Air Kerma Rate

Nuclide	Source ID	NPL Calibration Certificate	Air <u>Kerma Rate</u> ($\mu\text{Gy}\cdot\text{m}^2\cdot\text{h}^{-1}$)	Uncertainty	Reference Time
^{125}I	S1707/196 6 17126	E07100248/1	0.545	$\pm 5\%$	31 December 2007 1200 GMT
^{137}Cs	67505 E8-287	E07100248/2	38.2	$\pm 3\%$	1 February 2008 1200 GMT
^{192}Ir	IRF-1 (Ir2.A81) Batch 956	E07100248/3	59.3	$\pm 3\%$	3 January 2008 1200 GMT

Calibration of HDR 1000 Plus Well Chamber

Table 2 – Calibration factors N_{Sk} and N_A , used to convert induced current to Air Kerma Rate and ‘Apparent Activity’ respectively.

Nuclide	Source ID	Corrected current (pA)	N_{Sk} (Gy.m ² .h ⁻¹ .A ⁻¹)	N_A (GBq.A ⁻¹)
¹²⁵ I	S1707/1966 17126	-2.189	-2.489 x 10 ⁵ ± 5%	-7.251 x 10 ⁹ see note (a)
¹³⁷ Cs	67505 E8-287	-75.871	-5.035 x 10 ⁵ ± 3%	-6.484 x 10 ⁹ see note (b)
¹⁹² Ir	IRF-1 (Ir2.A81) Batch 956	-136.6	-4.341 x 10 ⁵ ± 3%	-3.983 x 10 ⁹ see note (c)

Ongoing Work :

- Derivation of Calibration factors for NPL Secondary Standard Ionisation Chamber
- Use NPL Holder, or design another ??
- I-125 Seeds
 - Existing factors for Oncoseed
 - IMC6711 Single Seed
 - IMC7000 Rapid Strand
 - Various Manufacturers
 - BEBIG : Seeds and Strands
 - ISOCORD S06
 - ISOSEED S17
 - ?

Y-90 Comparison Exercise

- Comparison for UK hospitals
 - Planned early 2009
- Aim to include manufacturers/suppliers
- NIST have recently performed a similar exercise
 - Ask Brian/Jeff ?

Commercial Calibrator Linearity