


NEW ELECTROMETER SYSTEM FOR WELL TYPE IONISATION CHAMBERS

John Sephton


ICRM Life Sciences Working Group Meeting

NPL


November 2008




ION CHAMBER SCHEMATIC



NPL AMPLIFIER BASED ELECTROMETER




SAMPLING WITH NPL BASED ELECTROMETER



KEITHLEY 6514 BASED ELECTROMETER

CAPACITOR
VOLTAGE



Voltage readings are corrected for decay during measurement

$$I = C \times dV/dT$$



Initial Estimate of Current	Electrometer Range	Upper limit of voltage
≥ 100 pA	20 V	19 V
≥ 1 pA and < 100 pA	2 V	2 V or voltage after 1000 readings
< 1 pA	2 V	2 V or voltage after 2000 readings

Electrometer settings used at various currents

TRACEABILITY

$$I = C \times dV/dT$$

VOLTAGE: Keithley electrometer calibrated with traceable voltage source

TIME: Pulse unit calibrated with traceable frequency meter

CAPACITANCE: Comparative bridge measurement made against standard capacitor

Type A uncertainty:

Random noise

Typically 2 fA

Type B uncertainty:

Feedback capacitor drift
and leakage

Voltmeter error

Current	Type A Uncertainty (%) (5 minute run)	Type B Uncertainty (%)
1 nA	0.0002	0.05
100 pA	0.002	0.03
10 pA	0.02	0.05
1 pA	0.2	0.5
100 fA	2	5

Typical Type A (statistical) and Type B (non-statistical) uncertainties at $k = 2$
(95% confidence limits)

Acknowledgements

Development:

Jonathan Williams

J T Janssen

Nick Fletcher (now at BIPM)

Software:

Keith Lines

Testing:

Stephen Giblin

Eleanor Bakhshandear

Stanislav Stanev (Bulgarian Institute Of
Metrology)

John Keightley

Andy Fenwick

Any questions?